

GAIN CHIPS AR coated Fabry-Perot Laser

Distributor

....

amstechnologies

where technologies meet solutions

info@amstechnologies.com www.amstechnologies-webshop.com

TOPTICA

Contact us 🕙

General Product Information

Product	Application
tunable 690 nm Fabry-Perot Laser	Spectroscopy
for use in an External Cavity Diode Laser (ECDL)	
sealed SOT Housing	
Monitor Diode	



Absolute Maximum Ratings

Parameter	Symbol	Unit	min	typ	max
Storage Temperature	Ts	°C	-40		85
Operational Temperature at Case	T _c	°C	0		30
Forward Current	١ _F	mA			160
Reverse Voltage	V _R	V			0
Output Power (extracavity)	Popt	mW			20

Recommended Operational Conditions

Parameter	Symbol	Unit	min	typ	max
Operational Temperature at Case	T _C	°C		20	
Forward Current	١ _F	mA			140

Characteristics ex-cavity at T_C= 20°C, BOL under recommended working condition

Parameter	Symbol	Unit	min	typ	max
Center Wavelength	λ_{C}	nm		685	
Tuning Range	$\Delta\lambda_{tun}$	nm	675		692
Output Power	P _{opt}	mW		15	
Polarization				TE	
Spatial Mode (transversal)				TEM ₀₀	
Monitor Detector Responsivity	I_{mon} / P_{ASE}	µA/mW	1		40

Measurement Conditions / Comments

Stess in excess of the Absolute Maximum Ratings can cause permanent damage to the device.

Measurement Conditions / Comments

Measurement Conditions / Comments The actual achieved wavelength and power are strongly influenced by the external cavity. eyP gives no guarantee on these parameters.

E field parallel to Pin 2 - Pin 3 - plane

Fundamental Mode

 $U_{R MD} = 5 V$

Revision 1.01

GAIN CHIPS AR coated Fabry-Perot Laser

Chip Parameter						
Parameter	Symbol	Unit	min	typ	max	Measurement Conditions / Commer
Cavity Length	L	μm		1000		
Reflectivity at Front Facet	R _{ff}			3·10 ⁻⁴	1·10 ⁻³	


Package Dimensions

Parameter	Symbol	Unit	min	typ	max
Height of Emission Plane	h	mm	3.50	3.65	3.70
Excentricity of Emission Center	R	mm			0.12
Pin Length	L _{PIN}	mm		14	

Package Pinout

- 1 Laser Diode Cathode, Monitor Diode Cathode, Case
- 2 Photo Diode Anode

3 Laser Diode Anode



Measurement Conditions / Comments reference plane: top side of TO header

reference: center of outer diameter of header




© All rights reserved by eagleyard Photonics GmbH. This data sheet will be electronically administered and is subject to change without notice. Uncontrolled copy when printed.



2019-03-01

Revision 1.01

GAIN CHIPS AR coated Fabry-Perot Laser





AIZ-16-0421-1517

© All rights reserved by eagleyard Photonics GmbH. This data sheet will be electronically administered and is subject to change without notice. Uncontrolled copy when printed.






Revision 1.01

GAIN CHIPS AR coated Fabry-Perot Laser

Typical Measurement Results





Performance figures, data and any illustrative material provided in this specification are typical and must be specifically confirmed in writing by eagleyard Photonics before they become applicable to any particular order or contract. In accordance with the eagleyard Photonics policy of continuous improvement specifications may change without notice.

Unpacking, Installation and Laser Safety

Unpacking the laser diodes should only be done at electrostatic safe workstations (EPA). Though protection against electro static discharge (ESD) is implemented in the laser package, charges may occur at surfaces. Please store this product in its original package at a dry, clean place until final use. During device installation, ESD protection has to be maintained.

The TPA diode type is known to be sensitive against thermal stress. It should not be operated without appropriate injection from a seed laser. Operating at moderate temperatures on proper heat sinks will contribute to a long lifetime of the diode. The chip should be protected against moisture. A water vapor content below 5000 ppm is recommended for applications with high reliability requirements.

The laser emission from this diode is close to the invisible infrared region of the electromagnetic spectrum. Avoid direct and/or indirect exposure to the free running beam. Collimating the free running beam with optics as common in optical instruments will increase threat to the human eye.

Each laser diode will come with an individual test protocol verifying the parameters given in this document.







2019-03-01